

**DEPARTMENT OF  
PUBLIC HEALTH AND HUMAN SERVICES**

**BRIAN SCHWEITZER  
GOVERNOR**

**JOAN MILES  
DIRECTOR**

**STATE OF MONTANA**

[www.dphhs.mt.gov](http://www.dphhs.mt.gov)

PO Box 4210  
HELENA, MT 59604-4210

APRIL 21, 2006

STATE OF MONTANA REQUEST FOR PROPOSAL ADDENDUM  
IFB #: 0610048GU  
TO BE OPENED: APRIL 26, 2006  
TITLE: HPLC INSTRUMENT

**ADDENDUM NO.2**

TO ALL OFFERORS:

The following question has been asked to clarify the specifications:

- Q. Is the post column module going to be used for Carbamates, Paraquats and Diquats in addition to Glyphosates?
- A. YES, the post column module going to be used for Carbamates, Paraquats and Diquats in addition to Glyphosates.

**Acknowledgement of Addendum:**

The offeror for this solicitation must acknowledge receipt of this addendum. This page must be submitted at the time set for the proposal opening or the proposal may be disqualified from further consideration.

I acknowledge receipt of Addendum #1.

Signed: \_\_\_\_\_

Company Name: \_\_\_\_\_

Date: \_\_\_\_\_

Sincerely,

Gwen Ungerman  
DPHHS Contracts Officer  
406-444-0546